



America

# CERTIFICATE

No. U8 16 12 19612 025

Holder of Certificate: **Amann Girrbach AG**

Herrschaftswiesen 1  
6842 Koblach  
AUSTRIA

Production

19612

Facility(ies):

Certification Mark:



**C US**  
**Laboratory equipment**

Product:

Model(s):

Ceramill Argotherm 2  
Aidite AF-CoCr  
Dental Max T2 CoCr  
IDC T2 CoCr  
Novux nx M-Therm  
Straumann Argotherm  
Ceramill Argotherm 2 JPN  
Ceramill Therm 3  
IDC T1 ZI  
Novux nx Z-Therm  
Straumann Therm  
Ceramill Therm 3 JPN

Parameters:

Rated voltage: 220-240V or 200V  
Rated frequency: 50-60Hz  
Rated power consumption: 3500W

Protection class: I

Tested according to:

UL 61010-1:2010/R:2016-04  
UL 61010-2-010:2015-01  
CAN/CSA-C22.2 No. 61010-1:2012 + UPD No. 1:2015-07  
+ UPD No. 2:2016-04  
CAN/CSA-C22.2 No. 61010-2-010:2015-01

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713084669-000

Date, 2016-12-09

Page 1 of 1

