



Ref. Certif. No.

DE 3 - 30751

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE**  
**CERTIFICAT D'ESSAI OC**

Product

Laboratory Equipment  
(dental scanner)

Name and address of the applicant

Amann Girrbach AG  
Herrschaftswiesen 1  
6842 Koblach, AUSTRIA

Name and address of the manufacturer

Amann Girrbach AG  
Herrschaftswiesen 1, 6842 Koblach, AUSTRIA

Name and address of the factory

smart optics Sensortechnik GmbH  
Lise-Meitner-Allee 10, 44801 Bochum, GERMANY

Ratings and principal characteristics

Rated voltage: 100-240V  
Rated frequency: 50/60Hz  
Rated power  
consumption: 80W  
Protection class: I

Trade mark (if any)

Amann Girrbach

Model/type Ref.

Ceramill MAP400  
Ceramill MAP400+  
IDC Scan  
HLD Scan  
Aidite Scan  
Novux nx 100

Additional information (if necessary)

Certificate DE 3 - 30284 issued 2016-08-22 is replaced by this  
version due to technical changesA sample of the product was tested and found  
to be in conformity with

IEC 61010-1:2010

as shown in the Test Report Ref. No.  
which forms part of this certificate

028-713014318-300

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date,

2018-07-10  
CB 18 07 19612 053

Erich Thurner



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service