


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (dental scanner)
Name and address of the applicant  AMANNGIRRBACH	Amann Girschbach AG Herrschaftswiesen 1 6842 Koblach AUSTRIA
Name and address of the manufacturer	Amann Girschbach AG Herrschaftswiesen 1, 6842 Koblach, AUSTRIA
Name and address of the factory	smart optics Sensortechnik GmbH Lise-Meitner-Allee 10, 44801 Bochum, GERMANY
Ratings and principal characteristics	Rated voltage: 100-240V Rated frequency: 50-60Hz Rated power consumption: 60W Protection class: I
Trade mark (if any)	Amann Girschbach
Model/type Ref.	Ceramill MAP600, Ceramill MAP600+, IDC S3
Additional information (if necessary)	Certificate DE 3 - 30808 issued 2018-10-10 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with as shown in the Test Report Ref. No. which forms part of this certificate	IEC 61010-1:2010 713201782

Page 1 of 1

This CB Test Certificate is issued by the National Certification Body

CB 019612 0071 Rev. 00

Date, 2021-02-23



(Benedikt Pulver)