



Ref. Certif. No.

DE 3 - 30748

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC**

Product	Laboratory equipment (dental milling machine)
Name and address of the applicant	Amann Girrbach AG Herrschaftswiesen 1 6842 Koblach, AUSTRIA
Name and address of the manufacturer	Amann Girrbach AG Herrschaftswiesen 1, 6842 Koblach, AUSTRIA
Name and address of the factory	Amann Girrbach AG Schweizerstrasse 96, 6830 Rankweil, AUSTRIA
Ratings and principal characteristics	Rated voltage: 100-230V Rated frequency: 50-60Hz Rated power consumption: 250W Protection class: I
Trade mark (if any)	Amann Girrbach
Model/type Ref.	Ceramill Mikro (4X), IDC Mikro, Novux nx Mill4, Aidite AM-D4 Ceramill Mikro (5X), IDC Mikro (5X), Straumann D series
Additional information (if necessary)	Certificate DE 3 - 30498 issued 2017-07-06 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010
as shown in the Test Report Ref. No. which forms part of this certificate	028-713110692-100

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Date, 2018-07-10
CB 18 07 19612 056

Erich Thurner



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service