



Ref. Certif. No.

DE 3 - 30750

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC**CB TEST CERTIFICATE  
CERTIFICAT D'ESSAI OC**

Product	Laboratory equipment (dental milling machine)
Name and address of the applicant	Amann Girrbach AG Herrschaftswiesen 1 6842 Koblach, AUSTRIA
Name and address of the manufacturer	Amann Girrbach AG Herrschaftswiesen 1, 6842 Koblach, AUSTRIA
Name and address of the factory	Amann Girrbach AG Schweizerstrasse 96, 6830 Rankweil, AUSTRIA
Ratings and principal characteristics	Rated voltage: 100-230V Rated frequency: 50-60Hz Rated power consumption: 250W Protection class: I
Trade mark (if any)	Amann Girrbach
Model/type Ref.	Ceramill Mikro ic, IDC Mikro ic, Straumann C series, Practice lab mill, Carestream CS3100, TROPHYCAM PRO
Additional information (if necessary)	Certificate DE 3 - 30632 issued 2018-02-15 is replaced by this version due to technical changes
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010
as shown in the Test Report Ref. No. which forms part of this certificate	028-713110733-200

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2018-07-10  
CB 18 07 19612 057

Erich Thurner



TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

Product Service