


IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Laboratory Equipment (dental oven)
Name and address of the applicant  AMANNGIRRBACH	Amann Girrbach AG Herrschaftswiesen 1 6842 Koblach AUSTRIA
Name and address of the manufacturer	Amann Girrbach AG Herrschaftswiesen 1, 6842 Koblach, AUSTRIA
Name and address of the factory	Amann Girrbach AG Schweizerstrasse 96, 6830 Rankweil, AUSTRIA
Ratings and principal characteristics	Rated voltage: 200-230 V Rated frequency: 50-60 Hz Rated power consumption: 3500 W Protection class: I
Trade mark (if any)	AMANNGIRRBACH
Model/type Ref.	Ceramill Therm DRS
A sample of the product was tested and found to be in conformity with	IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016 IEC 61010-2-010:2019
as shown in the Test Report Ref. No. which forms part of this certificate	028-713191569-000

This CB Test Certificate is issued by the National Certification Body

CB 019612 0069 Rev. 00
Date, 2020-09-24

(Benedikt Pulver)