




America

CERTIFICATE

No. U8 17 04 19612 031

Holder of Certificate: Amann Girrbach AGHerrschaftswiesen 1
6842 Koblach
AUSTRIA**Production Facility(ies):** 19612**Certification Mark:****Product:** Laboratory Equipment**Model(s):** Ceramill Motion 2, Ceramill Motion 2 (4X)**Parameters:**
Rated voltage: 100-230V
Rated frequency: 50-60Hz
Rated power consumption: 250W
Protection class: I**Tested according to:** UL 61010-1:2012-05
CAN/CSA-C22.2 No. 61010-1:2012-05

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 028-713006992-200**Date,** 2017-04-20

Page 1 of 1

