



America

# CERTIFICATE

No. U8 16 10 19612 021

**Holder of Certificate:** Amann Girrbach AG

Herrschaftswiesen 1  
6842 Koblach  
AUSTRIA

**Production Facility(ies):**

19612

**Certification Mark:**



C US

**Product:**

Laboratory equipment

**Model(s):**

Giroform Pinbohrgerät

**Parameters:**

Rated voltage:	100/115/230V
Rated frequency:	50/60Hz
Rated power consumption:	75W
Protection class:	I
Laser class:	2

**Tested according to:**

UL 61010-1:2012-05  
CAN/CSA-C22.2 No. 61010-1:2012-05  
CAN/CSA-E60825-1:2015-11

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in anyway. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

**Test report no.:**

028-713051039-000

**Date,** 2016-10-14

Page 1 of 1

