

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD  
**ZERTIFIKAT** ◆ **CERTIFICATE** ◆ 認證書 ◆ **CERTIFICADO** ◆ **CERTIFICAT**



America

# CERTIFICATE

No. U8 019612 0060 Rev. 00

**Holder of Certificate:** **Amann Girrbach AG**  
 Herrschaftswiesen 1  
 6842 Koblach  
 AUSTRIA

**Production Facility(ies):** 082881  
**Certification Mark:**



**Product:** **Laboratory equipment**  
 (dental scanner)

**Model(s):** **Ceramill MAP600, IDC S3**

**Parameters:**  
 Rated voltage: 100-240V  
 Rated frequency: 50-60Hz  
 Rated power consumption: 60W  
 Protection class: I  
 Laser class: 1

**Tested according to:** CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04  
 UL 61010-1:2012/R:2016-04

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

**Test report no.:** 028-713134717-100

**Date,** 2018-10-10

( Abdul Sabbagh )