



America

# CERTIFICATE


No. U8 019612 0072 Rev. 00

**Holder of Certificate:** **Amann Girrbaach AG**



Herrschaftswiesen 1  
6842 Koblach  
AUSTRIA

**Certification Mark:**



**C US**

**Product:** **Laboratory equipment  
(dental scanner)**

**Model(s):** **Ceramill MAP600, Ceramill MAP600+, IDC S3**

**Parameters:**  
Rated voltage: 100-240V  
Rated frequency: 50-60Hz  
Rated power consumption: 60W  
Protection class: I

**Tested according to:** UL 61010-1:2012/R:2016-04  
CAN/CSA-C22.2 No. 61010-1:2012/U2:2016-04

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 713201782

**Date,** 2021-02-24

( Benedikt Pulver )