



America

# CERTIFICATE

No. U8 019612 0070 Rev. 00


**Holder of Certificate:** Amann Girrbach AG



Herrschaftswiesen 1  
6842 Koblach  
AUSTRIA

**Production Facility(ies):** 097234

**Certification Mark:**



**Product:** Laboratory equipment (dental oven)

**Model(s):** Ceramill Therm DRS

**Parameters:**  
Rated voltage: 200-230 V  
Rated frequency: 50-60 Hz  
Rated power consumption: 3500 W  
Protection class: I

**Tested according to:** UL 61010-1:2012/R:2018-11  
UL 61010-2-010:2019  
CAN/CSA-C22.2 No. 61010-1:2012/A1:2018-11  
CAN/CSA-C22.2 No. 61010-2-010:2019

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

**Test report no.:** 028-713191569-000

**Date,** 2020-09-24

( Benedikt Pulver )